



- 1 The electrode input of the PREAMP is designed for "Combination" electrodes. Connect the BNC connector as shown above.
- 2 When automatic temperature compensation (A.T.C.) is not required, place a 301Ω (1%) fixed resistor across PREAMP J-BOX OUTPUT terminals 1 and 2 in lieu of the A.T.C. element.
- 3 Proceed to install wiring from preamp J-Box output terminal strip to the GREAT LAKES Model 672 input terminal strip . Recommend use of multi-conductor cable as provided by GREAT LAKES, Belden #8724 or equivalent.

PARTS LIST	
DESCRIPTION	P/N
PREAMP	APA-S546
JUNCTION BOX	APB-120
PS-120-G130-B4 = P/N for Preamp and Junction Box Assembly	

Broadley-James Corporation			
1714 South Lyon Street, Santa Ana, CA 92705 Tel: (714) 547-8061 Fax: (714) 547-8988			
TITLE: STANDARD PREAMP INTERFACE TO GREAT LAKES INSTRUMENTS MODEL 672			
DRAWN BY:	R. RAMUS	APPROVED BY:	
DATE:	10/30/91	DATE:	
REVISION:	A	DATE:	1-DEC-94
			DRAWING NUMBER D-50130